

Cholesterol FS

Application for serum and plasma samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel

Identification	
This method is usable for analysis:	Yes
Twin reaction:	No
Name:	CHOL
Shortcut:	
Reagent barcode reference:	024
Host reference:	

Technic	
Type:	End point
First reagent:[µL]	180
Blank reagent	Yes
Sensitive to light	
Second reagent:[µL]	
Blank reagent	
Sensitive to light	
Main wavelength:[nm]	508
Secondary wavelength:[nm]	700
Polychromatic factor:	1.000
1 st reading time [min:sec]	(-0.12)
Last reading time [min:sec]	10:00
Reaction way:	Increasing
Linear Kinetics	
Linearity: Maximum deviation [%]	
Fixed Time Kinetics	
Endpoint	
Prozone Limit [%]	

Reagents	
Decimals	
Units	

Sample	
Diluent	DIL A (NaCl)
Hemolysis:	
Agent [µL]	0 (no hemolysis)
Cleaner	
Sample [µL]	0
Technical limits	
Concentration technical limits-Lower	1
Concentration technical limits-Upper	750
SERUM	
Normal volume [µL]	2
Normal dilution (factor)	1
Below normal volume [µL]	5
Below normal dilution (factor)	1
Above normal volume [µL]	2
Above normal dilution (factor)	6
URIN	
Normal volume [µL]	2
Normal dilution (factor)	1
Below normal volume [µL]	5
Below normal dilution (factor)	1
Above normal volume [µL]	2
Above normal dilution (factor)	6
PLASMA	
Normal volume [µL]	2
Normal dilution (factor)	1
Below normal volume [µL]	5
Below normal dilution (factor)	1
Above normal volume [µL]	2
Above normal dilution (factor)	6
CSF	
Normal volume [µL]	2
Normal dilution (factor)	1
Below normal volume [µL]	5
Below normal dilution (factor)	1
Above normal volume [µL]	2
Above normal dilution (factor)	6
Whole blood	
Normal volume [µL]	2
Normal dilution (factor)	1
Below normal volume [µL]	5
Below normal dilution (factor)	1
Above normal volume [µL]	2
Above normal dilution (factor)	6

Results	
Decimals	0
Units	mg/dL
Correlation factor-Offset	0.000
Correlation factor-Slope	1.000

Range	
Gender	All
Age	
SERUM	>= <=200
URINE	
PLASMA	>= <=200
CSF	
Whole blood	
Gender	
Age	
SERUM	
URINE	
PLASMA	
CSF	
Whole blood	

Contaminants	
Contaminant 1	Please refer to r910 Carryover Pair Table
Wash with	
Cycle	
Volume [µL]	
Contaminant 2	
Wash with	
Cycle	
Volume [µL]	
Contaminant 3	
Wash with	
Cycle	
Volume [µL]	

Calibrators details	
Calibrator list	Concentration
Cal. 1/Blank	0
Cal. 2	*
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
	Max delta abs.
Cal. 1	0.003
Cal. 2	0.020
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
Drift limit [%]	
Calculations	
Model	
Degree	X

* Enter calibrator value