

Creatinine FS

Application for serum, plasma and urine samples (uncompensated method)

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel

Identification	
This method is usable for analysis:	Yes
Twin reaction:	No
Name:	CREA
Shortcut:	
Reagent barcode reference:	032
Host reference:	

Technic	
Type:	Linear kinetic
First reagent:[μ L]	160
Blank reagent	Yes
Sensitive to light	
Second reagent:[μ L]	40
Blank reagent	Yes
Sensitive to light	
Main wavelength:[nm]	508
Secondary wavelength:[nm]	570
Polychromatic factor:	1.000
1 st reading time [min:sec]	5:48
Last reading time [min:sec]	7:36
Reaction way:	Increasing
Linear Kinetics	0.2200
Linearity: Maximum deviation [%]	100
Fixed Time Kinetics	
Endpoint	
Prozone Limit [%]	

Reagents	
Decimals	
Units	

Sample	
Diluent	DIL A (NaCl)
Hemolysis:	
Agent [μ L]	0 (no hemolysis)
Cleaner	
Sample [μ L]	0
Technical limits	
Concentration technical limits-Lower	0.1
Concentration technical limits-Upper	15
SERUM	
Normal volume [μ L]	12
Normal dilution (factor)	1
Below normal volume [μ L]	24
Below normal dilution (factor)	1
Above normal volume [μ L]	2
Above normal dilution (factor)	1
URIN	
Normal volume [μ L]	6
Normal dilution (factor)	25
Below normal volume [μ L]	12
Below normal dilution (factor)	25
Above normal volume [μ L]	3
Above normal dilution (factor)	25
PLASMA	
Normal volume [μ L]	12
Normal dilution (factor)	1
Below normal volume [μ L]	24
Below normal dilution (factor)	1
Above normal volume [μ L]	2
Above normal dilution (factor)	1
CSF	
Normal volume [μ L]	12
Normal dilution (factor)	1
Below normal volume [μ L]	24
Below normal dilution (factor)	1
Above normal volume [μ L]	2
Above normal dilution (factor)	1
Whole blood	
Normal volume [μ L]	12
Normal dilution (factor)	1
Below normal volume [μ L]	24
Below normal dilution (factor)	1
Above normal volume [μ L]	2
Above normal dilution (factor)	1

Results	
Decimals	2
Units	mg/dL
Correlation factor-Offset	0.000
Correlation factor-Slope	1.000

Range	
Gender	Male
Age	
SERUM	$\geq 0.9 \leq 1.3$
URINE	
PLASMA	$\geq 0.9 \leq 1.3$
CSF	
Whole blood	
Gender	Female
Age	
SERUM	$\geq 0.6 \leq 1.1$
URINE	
PLASMA	$\geq 0.6 \leq 1.1$
CSF	
Whole blood	

Contaminants	
Please refer to r910 Carryover Pair Table	

Calibrators details	
Calibrator list	Concentration
Cal. 1/Blank	0
Cal. 2	*
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
	Max delta abs.
Cal. 1	0.002
Cal. 2	0.004
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
Drift limit [%]	0.8
Calculations	
Model	X
Degree	1

* Enter calibrator value